

Actim[®] Pancreatitis

Ruling out acute pancreatitis
in a timely manner.



Actim[®] Pancreatitis



Technical specifications

Test principle	Qualitative immunochromatographic dipstick
Target enzyme	Trypsinogen-2
Specimen type	Urine
Reading time	5 minutes
Interpretation	Visual
Cut-off	50 µg/l
Test component	1 dipstick
Number of tests/kit	10 tests (32731 ETAC) 20 tests (32732 ETAC)
Storage	+2...+25 °C
Sample storage	Samples should be tested as soon as possible after collection. Samples may be stored in refrigerator at +2...+8 °C for 24 hours.

Actim Pancreatitis Controls kit (32700 ETAC) for quality control

Kit components	1 negative control (32780 ETAC) 1 positive control, low (32781 ETAC) 1 positive control, high (32782 ETAC) Reconstitution Solution (30889 ETAC)
Storage	+2...+8 °C

Ordering information

Product	Product code
Actim Pancreatitis, 10 tests	32731 ETAC
Actim Pancreatitis, 20 tests	32732 ETAC
Actim Pancreatitis Controls kit	32700 ETAC

Actim Oy

– part of Medix Biochemica Group
Klovinpellontie 3, FI-02180 Espoo, Finland
actim@actimtest.com

www.actimtest.com

